

Fluoromount™ **Aqueous Mounting Medium for Fluorescent Staining**

- Catalog No.:** K 024-200
- Size:** 200mL
- Intended Use:** Mounting of tissue specimens stained with fluorescent dyes.
- Introduction:** Fluoromount is an aqueous based mounting medium for cover slipping tissues stained with fluorescent dyes. This product is compatible with FITC, phycoerythrin, phycocyanin, and allophycocyanin. It can also be used with other fluorescent markers such as Texas Red and rhodamine.
- Format:** 200mL ready to use clear viscous aqueous mounting medium.
- Storage:** Store at room temperature. Product contains Sodium Azide as a preservative. Do not use beyond the expiration date stated on the label.
- Mounting Procedure:**
1. Remove the red tip from the bottle and cut the tip off with a sharp razor or scissors. This will help achieve smooth flow of Fluoromount and prevent formation of tiny bubbles.
 2. Place the bottle upside down in a container before using. This will help clear Fluoromount of tiny bubbles.
 3. Blot excess water from the slide without letting tissue specimens dry. Make sure the tissue is wet prior to mounting.
 4. Apply 2-3 drops of Fluoromount on the tissue section and gently apply a cover slip. Allow to dry for 30-45 minutes at room temperature. Seal edges of the coverslip with nail polish for long term storage.
- Removal of Fluoromount:** Fluoromount is an aqueous-based mounting medium and can easily be removed by soaking slides in deionized water. Place slide in a beaker full of deionized water on a magnetic stirrer. Leave the slide stirring in water for few hours or overnight for complete removal of dry Fluoromount.

IVD: For In Vitro Diagnostic Use

DBS will not be held responsible for patent infringement or other violation that may occur with the use of our product

DBS

1020 Serpentine Lane, # 114, Pleasanton, CA 94566 Tel: 925 484 3350, Fax: 925 484 3390

Website: www.dbiosys.com e-mail: customersupport@dbiosys.com